

A novel deep learning-based approach for sleep apnea detection using single-lead ECG signals

Anh-Tu Nguyen^{*‡}, Thao Nguyen^{*‡}, Huy-Khiem Le^{*}, Huy-Hieu Pham^{*†}, and Cuong Do^{*†}

^{*} College of Engineering & Computer Science, VinUniversity, Hanoi, Vietnam

[†] VinUni-Illinois Smart Health Center, VinUniversity, Hanoi, Vietnam

[‡] These authors contributed equally to this work.

Abstract—Sleep apnea (SA) is a type of sleep disorder characterized by snoring and chronic sleeplessness, which can lead to serious conditions such as high blood pressure, heart failure, and cardiomyopathy (enlargement of the muscle tissue of the heart). The electrocardiogram (ECG) plays a critical role in identifying SA since it might reveal abnormal cardiac activity. Recent research on ECG-based SA detection has focused on feature engineering techniques that extract specific characteristics from multiple-lead ECG signals and use them as classification model inputs. In this study, a novel method of feature extraction which based on the detection of S peaks is proposed to enhance the detection of adjacent SA segments using a single-lead ECG. In particular, ECG features collected from a single lead (V2) are used to identify SA episodes. On the extracted features, a CNN model is trained to detect SA. Experimental results demonstrate that the proposed method detects SA from single-lead ECG data is more accurate than existing state-of-the-art methods, with 91.13% classification accuracy, 92.58% sensitivity, and 88.75% specificity. Moreover, the further usage of features associated with the S peaks enhances the classification accuracy by 0.85%. Our findings indicate that the proposed machine learning system has the potential to be an effective method for detecting SA episodes.

I. INTRODUCTION

Sleep apnea (SA) is the most prevalent breathing problem associated with sleep [1]. It causes people to repeatedly stop and start breathing during sleep. There are various types of sleep apnea, but obstructive sleep apnea is the most prominent, which occurs when upper airway muscles relax during sleep and obstruct the airway, preventing adequate airflow [1]. Patients' breathing may stop for 10 seconds or longer before the reflexes kick in and then resume breathing. Sleep apnea affects approximately 3% of normal-weight adults, but over 20% of obese individuals, and in general, men are more susceptible to sleep apnea than women [2].

Sleep apnea is frequently associated with heart disease and metabolic disorders such as diabetes [3]. Several studies have demonstrated a link between sleep apnea and health issues such as type 2 diabetes, strokes, heart attacks, and even a shorter lifespan [3]–[6]. It is essential to diagnose and detect sleep apnea in order to avoid long-term health repercussions. The questionnaire, which includes the STOP-Bang Questionnaire [7], [8] and Berline Questionnaire [9], is one method for screening patients at risk for SA syndrome. The gold standard for diagnosing sleep-related breathing disorders, however, is polysomnography — a method

for collecting physiologic parameters during sleep [10]. A polysomnogram (PSG) is a diagnostic test that uses electroencephalogram (EEG), electrooculogram (EOG), electromyogram (EMG), electrocardiogram (ECG), and pulse oximetry, in addition to airflow and respiratory effort, to identify the underlying reasons of sleep abnormalities [10]. However, this diagnostic procedure is time-consuming, expensive, and inconvenient. Patients must be connected to at least 22 electrodes for several nights in order to measure 11 channels of sleep signals in the laboratory using specialized equipment [11]. This is a barrier for patients to independently install and use the device at home. Moreover, to diagnose a patient, physicians must spend considerable time monitoring and interpreting that data. Consequently, the PSG-based technique is costly, complicated, and cumbersome. That trigger a need for a simple, affordable, and user-friendly alternative.

ECG is a viable tool for diagnosing and detecting sleep apnea, which has lately generated considerable interest. While some study groups have demonstrated that a patient's cardiovascular activity changes during sleep apnea, resulting in typical ECG readings [12], [13], other studies have demonstrated a link between patients with SA and other cardiovascular disorders [14]. Thus, the Apnea-ECG database [15] was developed by PhysioNet to spot abnormalities in patients' ECG signals when they have SA. Various research use ECG data to identify SA episodes, some of which employ traditional machine learning techniques based on feature engineering. [16]–[22] and others applying deep learning techniques [17], [23] for their classification tasks. In [24], Li *et al.* suggested a technique for detecting SA using ECG data, which is based on a deep neural network and a Hidden Markov model (HMM). The approach used a sparse auto-encoder to learn features, which were then fed into two types of classifiers (SVM and ANN). In the per-segment SA detection, they achieve 84.5% classification accuracy. In [22], a Sgolay filter was applied to extract the Heart Rate Variability (HRV) and the ECG-Derived Respiration (EDR), then they were used for the training procedure to achieve 82.2% accuracy. Shen *et al.* introduced a method that based on a weighted-loss time-dependent (WLTD) classification model and a multiscale dilation attention 1-D convolutional neural network (MSDA-1DCNN) [18]. That study obtained

89.4% accuracy, 89.8% sensitivity and 89.1% specificity. A novel method for diagnosing SA using a pre-trained AlexNet model is reported in [21], in which per-minute segments of a single-lead ECG recording are decomposed using continuous wavelet transform (CWT) and subsequently 2D scalogram images are created. Following that, a CNN based on a deep learning algorithm is used to improve classification performance.

While recent studies have sought to utilize ECG signals as direct input to deep learning models, research using hand-crafted features offers the potential for development because of its transparency and interpretability. However, recent research targeted at identifying SA by feature engineering only consider the position of the R peak in the ECG, neglecting the remaining four peaks (P, Q, S, and T). In this work, we not only identify the R peak, but also determine the position of the S peak and demonstrate that using the S peak enhances the model’s performance. To reduce noise and signal artifacts, we first applied a Finite Impulse Response (FIR) band-pass filter to the signal. Afterward, features are extracted based on the detection of the R and S peaks. We employ the feature extraction approach given by Wang *et al.* [25], in which an ECG record is divided into 5 minute-long segments, then the R peak is detected. In this work, the detection of S peak is added, followed by cubic interpolation to generate 900 values for each feature. The extracted features were used as input to the SE-ResNext 50 model to classify ECG signals with and without SA. Our model achieved 91.13% classification accuracy, 92.58% sensitivity, and 88.75% specificity. It has been demonstrated that using additional features extracted from S peaks improves accuracy by 0.85% compared to using only R-peak-related features, which may indicate anomalies in the S peak morphology during SA.

The rest of this paper is structured as follows. Methods for ECG signal preprocessing, feature extraction, and classification are introduced in section II. Section III describes the dataset and experimental setups. In Section III and V we report experimental results and summarize the key findings of this work.

II. METHODS

Methods for preprocessing the ECG signal, extracting hand-crafted features, and developing a classifier for ECG with and without SA are described in this section. The proposed system’s schematic is depicted in Fig. 1.

A. Pre-processing data

ECG signal is frequently contaminated by a variety of noise sources, such as 50/60 Hz interference from power lines, EMG signal from muscles, motion artefacts, and variations in electrode-skin contact. Therefore, a band-pass filter with a frequency range of 8 to 12 Hz was applied to remove noise and artifacts while maintaining the ECG signal’s QRS complex properties [26]. The signals before and after passing the band-pass filter are shown in Fig. 2a and Fig. 2b, respectively.





Fig. 1: An overview of the proposed approach for detecting sleep apnea.

B. Feature extraction


In order to detect ECG segments in patients with SA, abnormal characteristics of the ECG signal are extracted and fed into a machine learning classifier. As low QRS voltage and a leftward shift of the electrical axis may be associated with certain ECG abnormalities of SA patients [27], features associated to the QRS complex of the ECG signal can be utilized to detect SA episodes efficiently. Previous studies have classified ECG signals with and without SA based on characteristics associated to R peaks, which has limitations as other peaks are neglected (P, Q, S, T). To the best of our knowledge, this study is the first to recover features linked to S peaks and then then feed them into a CNN model in order to distinguish between SA and non-SA ECG. We use the following features for the classification purpose:

- Amplitude of R peaks
- Amplitude of S peaks
- RR interval (duration between two consecutive R peaks)
- SS interval (duration between consecutive S peaks)


The distributions of R-peak-related features and S-peak-related features are depicted in Fig. 3 and Fig. 4, respectively. It seems that the RR and SS intervals of SA ECG are the same as those of Non-SA ECG. It is noteworthy that R peaks in SA ECG have significantly lower amplitude than those in




(a)



(a)



(b)



(b)

Fig. 2: (a) The original signal of V2 lead ECG and (b) The component falls within the 8 to 20 Hz frequency range of the ECG signal.

Fig. 3: (a) The histogram of R peaks and (b) The histogram of RR intervals.

Non-SA ECG. In contrast, S peaks in SA ECG have a higher amplitude than those in non-SA. These analyses suggest that these features can be used to distinguish between SA and Non-SA.

To determine the positions of the R and S peaks, the following steps are taken:

- R peaks detection:** In order to determine the positions of the R peaks and calculate the RR intervals, we first used the Hamilton algorithm [28] to locate the R peaks. The amplitude of the R peaks are then extracted, and their positions are utilized to estimate the RR intervals. To remove redundant R peaks due to false detection, a local median filter proposed in [29] is applied. As the suspected irregular RR intervals can be caused by either false R peaks or missed R peaks, a lower bound and an upper bound are defined based on the physiological range of RR intervals in order to distinguish between these two types of uninterpretable data points. For abnormal RR intervals caused by false R peaks detection, the RR intervals in a sliding window are compared to the lower bound and


rectified using either averaging or merging procedures. For irregular RR intervals caused by a missed R wave detection, the RR intervals in the current window are divided into several equal values or averaged with the neighboring window based on the specified criteria.

- S peaks detection:** A method for detecting S peaks based on the position of R peaks is proposed, in which, the resulting S peak is the first negative peak (in the case of a positive R peak). S peak detection algorithm is specified in algorithm 1.


Fig. 5a and Fig. 5b display the location of the R and S peaks found by the algorithm described above on the filtered signal and the corresponding raw signal, respectively.

The effectiveness of the method for detecting R and S peaks is evaluated by manually labeling the positions of R and S peaks, followed by a comparison with the algorithm’s return values. We randomly labeled 200 R peaks and 200 S peaks in both classes (with and without SA); the accuracy and F1-score of R and S peaks detection algorithms are reported in Table I. All the statistics are performed on the *PhysioNet Apnea-ECG dataset*.

Due to the disparity in units (amplitude in millivolts and



(a)



(b)


Fig. 4: (a) The histogram of S peaks and (b) The histogram of SS intervals.

Algorithm 1 Algorithm for determining the ECG S peaks


```

function FIND_S_PEAKS(data_ecg, R_peaks)
    num_peaks ← R_peaks.shape[0]
    S_peaks ← list()
    N ← length(num_peaks)
    for index ← 0 to N do
        i ← R_peaks[index]
        cnt ← i
        if cnt+1 >= data_ecg.shape[0] then
            break
        end if
        while data_ecg[cnt] > data_ecg[cnt+1] do
            cnt ← cnt+1
            if cnt >= data_ecg.shape[0] then
                break
            end if
        end while
        S_peaks.append(cnt)
    end for
    return S_peaks
end function
    
```

time interval in seconds), feature values are normalized to the normal distribution before feeding them into the neural network. Based on the findings on study [25], cubic interpolation



(a)



(b)

Fig. 5: Positions R and S peaks are detected on the original signal (a) and on the filtered signal (b).

TABLE I: R and S peaks detection results for filtered signal.

	#beats	TP	TN	FP	FN	Acc	F1-score
R peaks	200	196	0	0	4	98%	98.99%
S peaks	200	196	0	0	4	98%	98.99%

yielding 900 points of each feature every 5-minute segment is utilized as an effective data augmentation technique.

C. Performance metrics

In this study, to evaluate the performance of the machine learning classifier, we adopt the accuracy, sensitivity, specificity and F1-score as evaluation metrics.

- **Sensitivity:** Sensitivity of a class is defined as the ratio of correctly classified samples to total number of samples actually belonging to that class.

$$Sensitivity = \frac{TP}{TP + FN} \quad (1)$$

- **Specificity:** Specificity is used to measure the proportion of negatives that are correctly identified. It is defined as the ratio of true negatives predicted to total number of

samples which belong to negative class.

$$Specificity = \frac{TN}{TN + FP} \quad (2)$$

- Accuracy: It is defined as the ratio of number of correctly classified samples to that of total samples.

$$Accuracy = \frac{TP + TN}{TP + FN + TN + FP} \quad (3)$$

- F1-score: It is the harmonic mean of the precision and recall.

$$F1\text{-score} = \frac{2TP}{2TP + FP + FN} \quad (4)$$

where TP : True Positive, TN : True Negative, FP : False Positive, FN : False Negative.

III. EXPERIMENTS

A. Dataset

The PhysioNet Apnea-ECG dataset used to study SA was made available by Philipps University [15]. The dataset consists of 70 single-lead ECG records (35 recordings from the public set and 35 from the withheld set). Each recording ranges in length from just under 7 hours to over 10 hours and consists of a digitized ECG signal and a set of apnea annotations, which is derived by human experts on the basis of simultaneously recorded respiration and related signals. The ECG signal is separated into 1-minute segments and labeled; a signal segment is identified as having SA if an apnea event occurs during that minute. Table II provides details on the number of 1-minute ECG signal segments identified with and without SA in the training and test sets.

TABLE II: Number of 1-minute ECG signal segments labeled with and without SA on training and test sets.

	Training set	Test set
SA	6,473 (38.74%)	6,490 (38.30%)
Non SA	10,236 (61.26%)	10,455 (61.79%)
Total	16,709	16,945

Previous research has demonstrated that adjacent segments provide valuable information for SA detection [25], [30]. We adopt the sampling method in [25], in which, each 1-minute signal segment and the surrounding 2-minute signal are used to form a 5-minute signal segment, which is then used for preprocessing and classification.

B. Implementation details and training methodology

After being preprocessed, the extracted features will be utilized to train the 1D CNN model. The NVIDIA GeForce RTX 3080 Ti GPU, 31 GB of RAM, and Intel Core i9-10900X processor operating at 3.70 GHz are used in all experiments. With a mini-batch size of 256, we train a specific model for 100 epochs, evaluating each model after every epoch. The optimal model for each training procedure will be the

checkpoint with the highest F1-score. To assess the impact of employing features associated with S peaks, we set up two experiments as follows:

- 1) Only R peak amplitude and RR intervals, which are features related to R peaks, are used.
- 2) Use a combination of features related to R and S peaks, including amplitudes and intervals of R and S peaks.

The ECG signal classifier is constructed using the SE-ResNext 50 [31] model as its backbone. The SE-ResNext 50 [32] model is a variant of ResNet 50 with the replacement of the identity connection with a Squeeze-and-Excitation block. It enables feature recalibration, allowing the network to learn how to use global information to selectively emphasize informative characteristics and suppress less helpful ones. This model has demonstrated its effectiveness with 1D data and ECG signals in particular [33].

IV. RESULTS AND DISCUSSION

A. Model performance

As previously noted, in order to test the efficacy of using S peak-related features, we conducted experiments with and without these features. Table III shows the classification results of the model employing solely R-related features versus the model employing all types of features. Our approach (RR intervals + R Amplitude) reports an accuracy of 90.28%, a specificity of 90.44%, and a sensitivity of 90.00%. In terms of F1-score, we report an F1-score of 86.85 (for SA class) and 92.30 for non SA class. According to the results, using features associated to the S peak enhances the F1-score by up to 1.5% in the SA class and 0.54% in the Non-SA class. That finding indicates that the characteristics associated with the S peak may represent an alteration in the ECG signal of individuals with SA and may contribute to the capacity to classify SA and Non-SA.

B. Comparison to state-of-the-art

Table IV show the comparison between the proposed method and current state-of-the-art approaches to SA detection. We show that our approach surpassed almost all of the competitors, with the largest disparity in accuracy is up to 9.1%. These results indicate the robustness of the propose method.

TABLE III: Experimental results using different features.

Results						
Feature combination	Classifier	Acc (%)	Spe (%)	Sen (%)	F1-score (%)	
					SA	Non SA
RR intervals R Amplitude	SE ResNext 50	90.28	90.44	90.00	86.85	92.30
RR intervals R Amplitude SS intervals S Amplitude	SE ResNext 50	91.13	92.58	88.75	88.35	92.84

TABLE IV: Comparison with state-of-the-art approaches.

Study	Comparison			
	Classifier	Accuracy	Sensitivity	Specificity
Li et al. [24]	Decision fusion	83.80%	88.90%	88.40%
André Pinho et al. [22]	ANN	82.12%	88.41%	89.10%
Mahsa Bahrami et al. [25]	LeNet-5	87.6%	83.10%	90.30%
Shen et al. [18]	IDCNN WLTD	89.40%	89.80%	89.10%
Mahsa Bahrami et al. [17]	ZFNet GRU	88.13%	84.26%	92.27%
Singh et al. [21]	Scalogram	86.22%	90.00%	83.82%
Kaicheng Feng et al. [16]	FSSAE	85.10%	86.20%	84.40%
Ours [a]	SE-ResNext 50	90.28%	90.44%	90.00%
Ours [b]	SE-ResNext 50	91.13%	92.58%	88.75%

a: The model solely uses features relating to R peaks (RR intervals and R amplitude).

b: The model uses additional S-peak-related features (RR intervals, R amplitude, SS intervals and S amplitude).

V. CONCLUSIONS

In this study, we introduce a novel method for detecting sleep apnea based on SS intervals and S amplitudes. Our experimental results demonstrate that the proposed technique is effective for SA detection, and its performance outperforms state-of-the-art works. Although showing a high-level of performance, our approach has several disadvantages. For example, the lack of access to SA patients’ ECG data and the large volume of data are the primary limitations of this study. Although proposed method has been tested with a number of experimental setups, it should be evaluated on a larger dataset.

In the future, we expect to integrate a SA detection module into a mobile application, which will alert patients suffering from SA to wake them up. Additionally, we also consider multi-model learning approaches [34], e.g. integrating demographic data and clinical data [35] to boost model performance. An example can be the image and encoded text (in the form of a graph neural network or a knowledge graph) fusion for better representation learning [36]–[39]. Not only that, we will use the explainable method to indicate which segments are at risk of disease and improve the accuracy of the model [40]. Specifically, we will highlight which segments of the ECG signal are suspect in the ECG signal obtained using an eXplainable. As a result, this application may save the patient’s life by restoring normal breathing and consciousness without complex setups.

VI. ACKNOWLEDGEMENT

The authors would like to thank the VinUni-Illinois Smart Health Center and the source of funding : VinUni Seed Grant 2020. For more information, please visit the website at <https://vinuni.edu.vn/vinuni-research/seed-funding-program/>. Beside, this work was also funded by Vingroup Joint Stock Company (Vingroup JSC), Vingroup, and supported by Vin-

group Innovation Foundation (VINIF) under project code VINIF.2021.DA00128.

REFERENCES

- [1] Shahrokh Javaheri, Ferran Barbe, Francisco Campos-Rodriguez, Jerome A Dempsey, Rami Khayat, Sogol Javaheri, Atul Malhotra, Miguel A Martinez-Garcia, Reena Mehra, Allan I Pack, et al. Sleep apnea: types, mechanisms, and clinical cardiovascular consequences. *Journal of the American College of Cardiology*, 69(7):841–858, 2017.
- [2] Johns Hopkins Medicine. The dangers of uncontrolled sleep apnea, 2021.
- [3] Richard ST Leung and T Douglas Bradley. Sleep apnea and cardiovascular disease. *American journal of respiratory and critical care medicine*, 164(12):2147–2165, 2001.
- [4] Terry Young, Mari Palta, Jerome Dempsey, Paul E Peppard, F Javier Nieto, and K Mae Hla. Burden of sleep apnea: rationale, design, and major findings of the wisconsin sleep cohort study. *WJM: official publication of the State Medical Society of Wisconsin*, 108(5):246, 2009.
- [5] Renee S Aronsohn, Harry Whitmore, Eve Van Cauter, and Esra Tasali. Impact of untreated obstructive sleep apnea on glucose control in type 2 diabetes. *American journal of respiratory and critical care medicine*, 181(5):507–513, 2010.
- [6] Christopher John Worsnop, Matthew Thomas Naughton, Colin Edwin Barter, Trefor Owen Morgan, Adrienne Ila Anderson, and Robert J Pierce. The prevalence of obstructive sleep apnea in hypertensives. *American journal of respiratory and critical care medicine*, 157:111–115, 1998.
- [7] Frances Chung, Hairil R Abdullah, and Pu Liao. Stop-bang questionnaire: a practical approach to screen for obstructive sleep apnea. *Chest*, 149(3):631–638, 2016.
- [8] Mahesh Nagappa, Pu Liao, Jean Wong, Dennis Auckley, Satya Krishna Ramachandran, Stavros Memtsoudis, Babak Mokhlesi, and Frances Chung. Validation of the stop-bang questionnaire as a screening tool for obstructive sleep apnea among different populations: a systematic review and meta-analysis. *PLoS one*, 10(12):e0143697, 2015.
- [9] Nikolaus C Netzer, Riccardo A Stoohs, Cordula M Netzer, Kathryn Clark, and Kingman P Strohl. Using the berlin questionnaire to identify patients at risk for the sleep apnea syndrome. *Annals of internal medicine*, 131(7):485–491, 1999.
- [10] Jessica Vensel Rundo and Ralph Downey III. Polysomnography. *Handbook of clinical neurology*, 160:381–392, 2019.
- [11] Manish Sharma, Shreyansh Agarwal, and U Rajendra Acharya. Application of an optimal class of antisymmetric wavelet filter banks for obstructive sleep apnea diagnosis using ecg signals. *Computers in biology and medicine*, 100:100–113, 2018.
- [12] Tom V Cloward, James M Walker, Robert J Farney, and Jeffrey L Anderson. Left ventricular hypertrophy is a common echocardiographic abnormality in severe obstructive sleep apnea and reverses with nasal continuous positive airway pressure. *Chest*, 124(2):594–601, 2003.
- [13] Krzysztof Narkiewicz, Nicola Montano, Chiara Cogliati, Philippe JH Van De Borne, Mark E Dyken, and Virend K Somers. Altered cardiovascular variability in obstructive sleep apnea. *Circulation*, 98(11):1071–1077, 1998.
- [14] Abu SM Shamsuzzaman, Bernard J Gersh, and Virend K Somers. Obstructive sleep apnea: implications for cardiac and vascular disease. *Jama*, 290(14):1906–1914, 2003.
- [15] Thomas Penzel, George B Moody, Roger G Mark, Ary L Goldberger, and J Hermann Peter. The apnea-ecg database. pages 255–258, 2000.
- [16] Kaicheng Feng, Hengji Qin, Shan Wu, Weifeng Pan, and Guanzheng Liu. A sleep apnea detection method based on unsupervised feature learning and single-lead electrocardiogram. *IEEE Transactions on Instrumentation and Measurement*, 70:1–12, 2020.
- [17] Mahsa Bahrami and Mohamad Forouzanfar. Sleep apnea detection from single-lead ecg: a comprehensive analysis of machine learning and deep learning algorithms. *IEEE Transactions on Instrumentation and Measurement*, 71:1–11, 2022.
- [18] Qi Shen, Hengji Qin, Keming Wei, and Guanzheng Liu. Multiscale deep neural network for obstructive sleep apnea detection using rr interval from single-lead ecg signal. *IEEE Transactions on Instrumentation and Measurement*, 70:1–13, 2021.
- [19] Nuno Pombo, Bruno MC Silva, André Miguel Pinho, and Nuno Garcia. Classifier precision analysis for sleep apnea detection using ecg signals. *Ieee Access*, 8:200477–200485, 2020.

- [20] Binish Fatimah, Pushpendra Singh, Amit Singhal, and Ram Bilas Pachori. Detection of apnea events from ecg segments using fourier decomposition method. *Biomedical Signal Processing and Control*, 61:102005, 2020.
- [21] Sinam Ajitkumar Singh and Swanirbhar Majumder. A novel approach osa detection using single-lead ecg scalogram based on deep neural network. *Journal of Mechanics in Medicine and Biology*, 19(04):1950026, 2019.
- [22] André Pinho, Nuno Pombo, Bruno MC Silva, Kouamana Bousson, and Nuno Garcia. Towards an accurate sleep apnea detection based on ecg signal: The quintessential of a wise feature selection. *Applied Soft Computing*, 83:105568, 2019.
- [23] Debangshu Dey, Sayanti Chaudhuri, and Sugata Munshi. Obstructive sleep apnoea detection using convolutional neural network based deep learning framework. *Biomedical engineering letters*, 8(1):95–100, 2018.
- [24] Kunyang Li, Weifeng Pan, Yifan Li, Qing Jiang, and Guanzheng Liu. A method to detect sleep apnea based on deep neural network and hidden markov model using single-lead ecg signal. *Neurocomputing*, 294:94–101, 2018.
- [25] Tao Wang, Changhua Lu, Guohao Shen, and Feng Hong. Sleep apnea detection from a single-lead ecg signal with automatic feature-extraction through a modified lenet-5 convolutional neural network. *PeerJ*, 7:e7731, 2019.
- [26] Mohamed Elgendi, Mirjam Jonkman, and Friso De Boer. Frequency bands effects on qrs detection. *Biosignals*, 2003:2002, 2010.
- [27] Ljuba Bacharova, Epameinondas Triantafyllou, Christos Vazaios, Ivana Tomeckova, Ivana Paramicova, and Ruzena Tkacova. The effect of obstructive sleep apnea on qrs complex morphology. *Journal of electrocardiology*, 48(2):164–170, 2015.
- [28] Pat Hamilton. Open source ecg analysis. In *Computers in cardiology*, pages 101–104. IEEE, 2002.
- [29] Lili Chen, Xi Zhang, and Changyue Song. An automatic screening approach for obstructive sleep apnea diagnosis based on single-lead electrocardiogram. *IEEE transactions on automation science and engineering*, 12(1):106–115, 2014.
- [30] Azadeh Yadollahi and Zahra Moussavi. Acoustic obstructive sleep apnea detection. In *2009 annual international conference of the IEEE engineering in medicine and biology society*, pages 7110–7113. IEEE, 2009.
- [31] Jie Hu, Li Shen, and Gang Sun. Squeeze-and-excitation networks. In *Proceedings of the IEEE conference on computer vision and pattern recognition*, pages 7132–7141, 2018.
- [32] Kaiming He, Xiangyu Zhang, Shaoqing Ren, and Jian Sun. Deep residual learning for image recognition. In *Proceedings of the IEEE conference on computer vision and pattern recognition*, pages 770–778, 2016.
- [33] Dongqi Wang, Qinghua Meng, Dongming Chen, Hupo Zhang, and Lisheng Xu. Automatic detection of arrhythmia based on multi-resolution representation of ecg signal. *Sensors*, 20(6):1579, 2020.
- [34] Christian Moro, Jessica Smith, and Zane Stromberga. Multimodal learning in health sciences and medicine: Merging technologies to enhance student learning and communication. In *Biomedical visualisation*, pages 71–78. Springer, 2019.
- [35] Khiem H Le, Hieu H Pham, Thao BT Nguyen, Tu A Nguyen, and Cuong D Do. Enhancing deep learning-based 3-lead ecg classification with heartbeat counting and demographic data integration. *arXiv preprint arXiv:2208.07088*, 2022.
- [36] Wei Tan, Prayag Tiwari, Hari Mohan Pandey, Catarina Moreira, and Amit Kumar Jaiswal. Multimodal medical image fusion algorithm in the era of big data. *Neural Computing and Applications*, pages 1–21, 2020.
- [37] Feiran Huang, Xiaoming Zhang, Zhonghua Zhao, Jie Xu, and Zhoujun Li. Image-text sentiment analysis via deep multimodal attentive fusion. *Knowledge-Based Systems*, 167:26–37, 2019.
- [38] Trung Thanh Nguyen, Hoang Dang Nguyen, Thanh Hung Nguyen, Huy Hieu Pham, Ichiro Ide, and Phi Le Nguyen. A novel approach for pill-prescription matching with gnn assistance and contrastive learning. *arXiv preprint arXiv:2209.01152*, 2022.
- [39] Anh Duy Nguyen, Thuy Dung Nguyen, Huy Hieu Pham, Thanh Hung Nguyen, and Phi Le Nguyen. Image-based contextual pill recognition with medical knowledge graph assistance. *arXiv preprint arXiv:2208.02432*, 2022.
- [40] Khiem H. Le, Hieu H. Pham, Thao BT. Nguyen, Tu A. Nguyen, Tien N. Thanh, and Cuong D. Do. Lightx3ecg: A lightweight and explainable deep learning system for 3-lead electrocardiogram classification, 2022.